

## Re-Strainer (In-Line-Strainer) 100 µm – Sterile

### Specification

Catalog No.:	43-75100-40
Strainer color	yellow-transparent
Lot No. / EXP	on bag, top
Quantity	5 x 4 devices, sterile (radiation)
Mesh material	PET precision woven mono-filaments
Housing material	PP
Screw cap	Luer-lock-port for syringe or tube connection
Outlet	Luer-lock-port
Stability	Isopropanol, cell lysis buffers
Storage	Room Temperature



### Component Description

The Re-Strainer is a filtration device for high volumes of liquid to exclude or concentrate particles from liquids with the possibility of particle recovery. The 6 ml recovery reservoir can be accessed by unscrewing the cap. The Re-Strainer has two female luer-lock-ports.

These ports can be used to connect several Re-Strainer to a filter cascade (male to male luer-lock connector) or integrate the Re-Strainer as an in-line-strainer via luer-lock tube adaptor

The mesh is an integrated part of the housing to give the device maximum mechanical stability.

### Warning and Limitations

**The product is for research and development only, not for diagnostic or therapeutic use.**

---

Pluriselect Life Science worldwide  
Deutscher Platz 5c, D-04103 Leipzig, Germany  
Phone: +49 341 333858-0  
[info@pluriselect.com](mailto:info@pluriselect.com)

Pluriselect-usa, Inc. – North America  
Gillespie Field Business Park  
2030 Gillespie Way 103, El Cajon, CA, USA  
Phone: +1 619 202 4297  
[service@pluriselect-usa.com](mailto:service@pluriselect-usa.com)